

**BID # CON25-03**

**ADDENDUM NUMBER 01**

REQUEST FOR PROPOSALS FOR:

**Ceiling-Mounted DR Manual X-Ray System**

**7/18/2025**

**TO: ALL CONTRACT BIDDERS OF RECORD**

This Addendum forms a part of the Contract Documents and modifies the Request for Proposals dated July 14, 2025, with amendments and additions noted below. This Addendum consists of a total of two (2) pages.

**OFFER**

Amendment #1: The OFFER section is hereby expanded to include the below component of the X-Ray System:

(please include in your proposal)

**X-Ray Generator**

Phase 208v 50kw High Speed Generator

3 Phase High Frequency Generator SHF525 High Speed

High Frequency 25KHz, 1 tube operation

3 phase 208 VAC

Output power:

640 mA @ 78kVp

500 mA @ 100kVp

400 mA @ 125 kVp

320 mA @ 150 kVp

Automatic line compensation +/-10%

Microprocessor controlled with auto diagnostic and error code indication for easy maintenance

X-ray tube overload protection

X-ray tube H.U. available indication and continuous monitoring for x-ray tube protection

Control of x-ray tube number of exposures

Hand switch for preparation and exposure control

Light and sound indication for x-ray exposure

kVp Radiographic range, from 40 to 125 kVp in 1 kVp step

mA Radiographic range from 10 to 640mA in 19 step,

Renard scale 10, 12.5, 16, 20, 25, 32, 40, 50, 64, 80, 100, 125, 160, 200, 250, 320, 400,

500 and 640 mA

mAs range from 0.1 to 500 mAs in 38 steps, Renard scale

Exposure time range from 0.001 to 10 seconds

All other aspects of the RFP are to remain the same



**Acknowledge receipt of this Addendum with the Proposal Form.  
Failure to do so may disqualify the Bidder.**

NOTE: IN THE EVENT THAT YOUR BID HAS BEEN SENT TO THIS OFFICE PRIOR TO RECEIVING THIS ADDENDUM, RETURN THE ADDENDUM WITHIN THE SPECIFIED TIME WITH ANY CHANGES YOU MAY WISH TO MAKE AND MARK ON THE REMITTANCE ENVELOPE BID INVITATION NUMBER AND OPENING DATE. RETURNED ADDENDA WILL SUPERSEDE PREVIOUSLY SUBMITTED BID.

Bidder \_\_\_\_\_

By \_\_\_\_\_ Date \_\_\_\_\_  
(This Document Must Be Signed)

Name \_\_\_\_\_  
(Please Print or Type Name)